

Dr Karen Mullinger PhD

Joint Nottingham and Birmingham Lecturer and Anne McLaren Research Fellow

[School of Psychology \(/schools/psychology/index.aspx\)](/schools/psychology/index.aspx)

Contact details

Telephone [+44 \(0\)121 414 7615](tel:+441214147615) (tel: [+44 121 414 7615](tel:+441214147615))

Email k.mullinger@bham.ac.uk (<mailto:k.mullinger@bham.ac.uk>)

School of Psychology
University of Birmingham
Edgbaston
Birmingham
B15 2TT
UK



About

Dr Mullinger is an expert in simultaneous EEG-fMRI. She focuses on understanding the sources of the EEG artefacts and how to reduce them at source. Using the best EEG-fMRI techniques available she investigates neurovascular coupling to better understand brain function.

See [Dr Mullinger's Nottingham profile \(http://www.nottingham.ac.uk/physics/people/karen.mullinger\)](http://www.nottingham.ac.uk/physics/people/karen.mullinger)

Qualifications

- **PhD** (2005-2008) Physics and Astronomy
- **BSc** (2001-2004) Physics with Medical Physics
Degree Classification: First Class with Honours

Biography

Dr Mullinger has gained unique and wide-ranging experience in the development and application of simultaneous electroencephalography (EEG) and functional magnetic resonance imaging (fMRI) over the past 7 years whilst working with the first MR-compatible EEG system (from Brain Products, Germany) at the SPMMRC, Nottingham. This has involved collaboration with academic and industrial research leaders (e.g. Brain Products and Philips Medical Systems).

Dr Mullinger undertook a BSc in Physics with Medical Physics at the University of Nottingham and went on to complete her PhD at the Sir Peter Mansfield Magnetic Resonance Centre, University of Nottingham. Her PhD addressed many aspects of simultaneous EEG-fMRI recordings including: improving EEG artefact correction, reducing artefacts at ultra-high field to make simultaneous recordings feasible, and identifying the major sources of MR image distortion due to EEG hardware. Subsequently, as a Sir Peter Mansfield Research Fellow and Anne McLaren Research Fellow, she has played a leading role in research aimed at understanding the origins of the gradient and pulse artefacts in EEG data, as well as in the development of beamformer techniques for localising EEG sources in EEG-fMRI experiments. She has established a number of collaborations with academic and industrial research leaders.

The success of her work is reflected not only in her publication record, but also in three personal invitations to contribute book chapters. Additionally she has been invited to speak at several international conferences. The importance of her contribution in advancing EEG-fMRI research was recognised by the recent award of the prestigious International Society of Magnetic Resonance in Medicine Junior Fellowship (May, 2011).

She has recently taken a joint Lecturer post between the University of Birmingham (Psychology) and the University of Nottingham (Physics). She will use this unique post to develop her research interests and facilitate collaborations between the universities.

Postgraduate supervision

Currently she supervises a number of postgraduates at the University of Nottingham.

If you would like to contact her about her research and supervision please email or call her.

Research

Research interests

Dr Mullinger's research interests are focused on simultaneous EEG-fMRI. Combining these two techniques is highly desirable since they provide complementary information regarding brain function. She is interested in both the technical aspects of combining these two techniques as well as using them to investigate neurovascular coupling to gain a better understanding of brain function.

Technical aspects of EEG-fMRI

Performing EEG and fMRI together is very challenging as the MRI environment is extremely hostile for measuring small electrical signals from neurons using EEG. Whilst simultaneous EEG-fMRI has been available as a research tool for over a decade there are still significant limitations in the areas of brain function which can be researched with this neuroimaging tool due to the artefacts in the EEG signals. Dr Mullinger therefore investigates the sources of these artefacts. She uses this knowledge to develop methods to reduce the problematic artefacts through better experimental set-up and new EEG hardware design.

Neurovascular Coupling

BOLD fMRI is non-invasive and is the most widely used technique to study brain function in humans. However, the BOLD signal is caused by an interaction of changes in blood flow, volume and oxygenation related to changes in neuronal activity. As a result this complex signal is often difficult to interpret and the origins of changes in the BOLD signal can be unclear. Dr Mullinger uses simultaneous EEG-fMRI to measure EEG, BOLD and blood flow changes related to a stimulus to interrogate the origin of aspects of the BOLD response which are currently poorly understood. Dr Mullinger believes a better understanding of the BOLD signal this will open new avenues of

Publications

- Bowtell, R., Mullinger, K. J., Yan, W. X., Brookes, M. J. and Geirsdottir, G. B. (2009). "Exploiting a Better Understanding of the Artifacts in Integrated EEG and fMRI." *Psychophysiology* 46: S5-S5.
- Brookes, M. J., Mullinger, K. J., Stevenson, C. M., Morris, P. G. and Bowtell, R. (2008). "Simultaneous EEG source localisation and artifact rejection during concurrent fMRI by means of spatial filtering." *Neuroimage* 40(3): 1090-1104.
- Brookes, M. J., Vrba, J., Mullinger, K. J., Geirsdottir, G. B., Yan, W. X., Stevenson, C. M., Bowtell, R. and Morris, P. G. (2009). "Source localisation in concurrent EEG/fMRI: Applications at 7T." *Neuroimage* 45(2): 440-452.
- Chowdhury, M. E., Mullinger, K. J., Glover, P. and Bowtell, R. (2013). "Reference layer artefact subtraction (RLAS): A novel method of minimizing EEG artefacts during simultaneous fMRI." *Neuroimage*.
- Debener, S., Mullinger, K. J., Mazy, R. K. and Bowtell, R. W. (2008). "Properties of the ballistocardiogram artefact as revealed by EEG recordings at 1.5, 3 and 7 T static magnetic field strength." *International Journal of Psychophysiology* 67(3): 189-199.
- Jansen, M., White, T. P., Mullinger, K. J., Liddle, E. B., Gowland, P. A., Francis, S. T., Bowtell, R. and Liddle, P. F. (2012). "Motion-related artefacts in EEG predict neuronally plausible patterns of activation in fMRI data." *Neuroimage* 59(1): 261-270.
- Mullinger, K., Brookes, M., Stevenson, C., Morgan, P. and Bowtell, R. (2008a). "Exploring the feasibility of simultaneous electroencephalography/functional magnetic resonance imaging at 7 T." *Magnetic Resonance Imaging* 26(7): 968-977.
- Mullinger, K., Debener, S., Coxon, R. and Bowtell, R. (2008b). "Effects of simultaneous EEG recording on MRI data quality at 1.5, 3 and 7 tesla." *International Journal of Psychophysiology* 67(3): 178-188.
- Mullinger, K. J., Castellone, P. and Bowtell, R. (2013a). "Best current practice for obtaining high quality EEG data during simultaneous fMRI." *J Vis Exp*(76).
- Mullinger, K. J., Havenhand, J. and Bowtell, R. (2013b). "Identifying the sources of the pulse artefact in EEG recordings made inside an MR scanner." *Neuroimage* 71: 75-83.
- Mullinger, K. J., Mayhew, S. D., Bagshaw, A. P., Bowtell, R. and Francis, S. T. (2013c). "Poststimulus undershoots in cerebral blood flow and BOLD fMRI responses are modulated by poststimulus neuronal activity." *Proc Natl Acad Sci U S A* 110(33): 13636-13641.
- Mullinger, K. J., Morgan, P. S. and Bowtell, R. W. (2008c). "Improved artifact correction for combined electroencephalography/functional MRI by means of synchronization and use of vectorcardiogram recordings." *Journal of Magnetic Resonance Imaging* 27(3): 607-616.
- Mullinger, K. J., Yan, W. X. and Bowtell, R. (2011). "Reducing the gradient artefact in simultaneous EEG-fMRI by adjusting the subject's axial position." *Neuroimage* 54(3): 1942-1950.
- White, T. P., Jansen, M., Doege, K., Mullinger, K. J., Park, S. B., Liddle, E. B., Gowland, P. A., Francis, S. T., Bowtell, R. and Liddle, P. F. (2013). "Theta power during encoding predicts subsequent-memory performance and default mode network deactivation." *Human Brain Mapping* 34(11): 2929-2943.
- Yan, W. X., Mullinger, K. J., Brookes, M. J. and Bowtell, R. (2009). "Understanding gradient artefacts in simultaneous EEG/fMRI." *Neuroimage* 46(2): 459-471.
- Yan, W. X., Mullinger, K. J., Geirsdottir, G. B. and Bowtell, R. (2010). "Physical Modeling of Pulse Artefact Sources in Simultaneous EEG/fMRI." *Human Brain Mapping* 31(4): 604-620.

