





Migration Pathways: Project Update




Cassandra Rauert
ESR2

Micro emission test chambers



UNIVERSITY OF
BIRMINGHAM

Emission Chamber Experiments



UNIVERSITY OF
BIRMINGHAM

Emission Chamber Experiments

- *Volatiles*: Mass balance experiments
 - Filter paper spiked with standards
 - Run low air flow pump through chamber
 - Analyse filter paper, PUFs, chamber wall washes separately after use
 - Sum up %Recovery

Emission Chamber Experiments

□ Influence of Temperature – 1 week experiments

Room Temperature (21 °C)

<i>BFR</i>	<i>BDE-47</i>	<i>BDE-85</i>	<i>BDE-99</i>	<i>BDE-100</i>	<i>BDE-153</i>	<i>BDE-154</i>	<i>BDE-183</i>	<i>BDE-209</i>		<i>α-HBCD</i>	<i>β-HBCD</i>	<i>γ-HBCD</i>
<i>% on exit PUF</i>	ND	ND	ND	ND	ND	ND	ND	ND		0.8	0.6	0.6
<i>Total %Rec</i>	76	93	89	82	86	76	89	101		95	85	92

Heated Chamber (60 °C)

<i>BFR</i>	<i>BDE-47</i>	<i>BDE-85</i>	<i>BDE-99</i>	<i>BDE-100</i>	<i>BDE-153</i>	<i>BDE-154</i>	<i>BDE-183</i>	<i>BDE-209</i>		<i>α-HBCD</i>	<i>β-HBCD</i>	<i>γ-HBCD</i>
<i>% on exit PUF</i>	63	17	36	57	5	13	0.7	0.9		1.7	0.6	1.0
<i>Total %Rec</i>	63	42	53	66	45	41	71	114		86	53	41

Emission Chamber Experiments

□ Sink Effects (% Recovery)

<i>BFR</i>	<i>BDE-47</i>	<i>BDE-85</i>	<i>BDE-99</i>	<i>BDE-100</i>	<i>BDE-153</i>	<i>BDE-154</i>	<i>BDE-183</i>	<i>BDE-209</i>		<i>α-HBCD</i>	<i>β-HBCD</i>	<i>γ-HBCD</i>
<i>Filter Paper</i>	5.3	21	13	10	59	35	71	85		94	65	74
<i>Wash solvent 1</i>	38	32	41	41	30	43	5.6	4.3		2.2	1.5	1.1
<i>Wash solvent 2</i>	3.5	6.3	5.8	5.1	5.9	9.6	1.2	0		1.2	0	0
<i>Wash solvent 3</i>	2.0	2.2	2.5	2.4	2.5	4.8	0.5	0		0	0	0
<i>Total %Rec</i>	49	61	62	59	97	93	78	91		97	67	75

Emission Chamber Experiments

□ ‘Tweaking’ system to improve Recoveries

	<i>BFR</i>	<i>BDE-47</i>	<i>BDE-85</i>	<i>BDE-99</i>	<i>BDE-100</i>	<i>BDE-153</i>	<i>BDE-154</i>	<i>BDE-183</i>	<i>BDE-209</i>		<i>α-HBCD</i>	<i>β-HBCD</i>	<i>γ-HBCD</i>
<i>Room Temp</i>	<i>% on exit PUF</i>	ND	ND	ND	ND	ND	ND	ND	ND		0.8	0.6	0.6
	<i>%Rec</i>	76	93	89	82	86	76	89	101		95	85	92
<i>Heat 24 hrs</i>	<i>% on exit PUF</i>	10	1.3	3.5	6.8	0.3	1.4	0.4	0.3		0.7	0.3	0.4
	<i>%Rec</i>	30	82	71	60	76	78	86	88		105	40	39
<i>Chamber left equilibrate</i>	<i>% on exit PUF</i>	21	2.0	10	19	0.4	1.5	0.5	0.5		1.7	0.6	1.0
	<i>%Rec</i>	73	67	66	70	73	67	87	87		86	53	41
<i>Heat 1 week</i>	<i>% on exit PUF</i>	63	17	36	57	5.2	13	0.7	0.9		0.3	0.1	0.2
	<i>% on 80°C PUF</i>	ND	3.8	2.1	1.6	3.1	4.7	0.6	1.8		ND	ND	ND
	<i>%Rec</i>	63	42	53	66	45	41	71	114		59	9.3	9.9

Investigations of SRM 2585

- Old SRM batch – August 2009
 - Inconsistent results
- New SRM 2585 batch – February 2012
 - More reliable results
 - Possibility old SRM is past its used by date
- ‘In-house’ dust reference material prepared for further quality assurance measures

Investigations of SRM 2585

□ Results (ng/g)

	BFR	BDE-47	BDE-85	BDE-99	BDE-100	BDE-153	BDE-154	BDE-183	BDE-209		α -HBCD	β -HBCD	γ -HBCD
Old SRM 2585	Average (ng/mL)	455	47	991	170	129	82	51	2811		28	9	173
	Range	± 326	± 25	± 247	± 65	± 40	± 34	± 16	± 923		± 6	± 4	± 215
New SRM 2585	Average (ng/mL)	475	42	786	134	106	74	41	2142		18	7	118
	Range	± 25	± 3	± 41	± 9	± 17	± 4	± 5	± 401		± 4	± 0.6	± 28
Certified/ Indicative Values		498 ± 46	43.8 ± 1.6	892 ± 53	145 ± 11	119 ± 1	83.5 ± 2.0	43.0 ± 3.5	2510 ± 190		19 ± 3.7	4.3 ± 1.1	120 ± 22
In-house Ref Material	Average (ng/mL)	108	46	71	95	101	46	55	2386		277	120	1508
	Range	± 28	± 16	± 23	± 29	± 22	± 14	± 48	± 2174		± 51	± 18	± 2269


Presentations

- Abstract accepted for Marie Curie symposium “SCIENCE – Passion, Mission, Responsibilities”, Warsaw, September 2011
- Poster presentation at 6th Annual Conference of the POPs network, Birmingham, April 2012

Future work

- Investigations into HBCD results (thermal degradation??)
- Repeat experiments – consistent results
- Analyse real samples in chamber for volatiles profile (PUF furniture, carpets, etc)
- Secondment to VITO – September 2012

Thankyou



UNIVERSITY OF
BIRMINGHAM